

Certificate of Compliance

Certificate: 2137693 Master Contract: 159754

Project: 2137693 **Date Issued:** 2009/07/20

Issued to: Pepperl+Fuchs, Incorporated

1600 Enterprise Pky Twinsburg, OH 44087

USA

Attention: Greg Prosser

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Issued by: Oliver Johnson

Authorized Nick Alfano, Operations Manager **by:**

R. A.D.

PRODUCTS

CLASS 2252 85 - PROCESS CONTROL EQUIPMENT - Certified to US Standards

CLASS 2252 05 - PROCESS CONTROL EQUIPMENT

Frequency to Voltage/Current Converter, Model KFU8-FSSP-1.D, permanently connected, rated 115/230V, 47-63Hz, 5VA; 24Vdc, 5W.

Rotational Speed Monitor, Model KFU8-DW-1.D, permanently connected, rated 115/230V, 47-63Hz, 5VA; 24Vdc, 5W.



Certificate: 2137693 Master Contract: 159754

Project: 2137693 **Date Issued:** 2009/07/20

Max. Ambient: 50°C.

1. The above model is rated Pollution Degree 2, and Installation Category II.

2. Certified as components for use in other equipment where the suitability of the combination is to be determined by CSA International.

3. A suitable fire enclosure is to be provided in the end use.

4. The Determination of Accessible Parts (Cl. 6.2) is to be conducted in the end-use.

5. The subject models have been evaluated for Basic Insulation Only.

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

UL Std. No. 61010-1 (2nd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements